



The Challenges of Adopting a Wireless Nurse call system

¹Sarafadeen Leye Lawal, ²Aliyu Surajo, ³Livingston Sunday Aduku, ⁴Muhammad Ahmad Baballe*

^{1, 2, 3, 4}Department of Mechatronics Engineering, Nigerian Defence Academy (NDA), Kaduna, Nigeria.

DOI: [10.5281/zenodo.15388093](https://doi.org/10.5281/zenodo.15388093)

Submission Date: 10 April 2025 | Published Date: 12 May 2025

*Corresponding author: [Muhammad Ahmad Baballe](mailto:muhammad.ahmad.baballe@gmail.com)

Department of Mechatronics Engineering, Nigerian Defence Academy (NDA), Kaduna, Nigeria.

<https://orcid.org/0000-0001-9441-7023>

Abstract

The nurse caller device is used as a specialized communication tool between patients and the healthcare provider within the hospital to expedite the doctor's or nurse's response time in order to give patients with speedy care. Using the wireless nurse caller device improves the efficiency and organization of the two parties' communication. The Bluetooth module MH-10 is connected to the ATmega8 microcontroller, which serves as both the transmitter and the receiver. The data is processed by an ATmega8 microcontroller, which also creates characters on the LCD, turns it on, and sounds the buzzer to call for the doctor or nurse. The benefits, drawbacks, and difficulties of implementing the nurse wireless calling system have all been discussed in this study.

Keywords: Internet of Things (IoT), Challenges, Solutions, Hospital, Nurse Caller, Impacts.

I. INTRODUCTION

The hospital is an essential part of a social and health organization with the function of providing plenary (comprehensive), healing (curative), and disease prevention (preventive) services to the community, whose services are handled by nurses, doctors, and other health experts. The nurses are considered the spearhead in the organization of hospital services because it is the nurse who continually provides care to the patients. The nurse continues to monitor the patient's health and provide care to them. Numerous previous studies have conducted research on nurse callers. The techno-economic evaluation of an ontology-based nurse call system through discrete event simulations was investigated by Vannieuwenborg [1]. The design and implementation of the Wired Nurse Call System was researched by Sharma [2]. A study on the automatic evaluation method of designating and invoking nurse education was researched by Maekawa [3]. Khara [4] researched the development of smart house call systems and android-based nurses for different abilities. A real-time feedback-centric nurse call system with archival monitoring using a Raspberry Pi was studied by Mahmud [5]. The hospital is equipped with a calling device in each patient's bedroom to speed up patient services. Presently, the nurse calling device used is very conventional, namely, it still uses a pair of cables. Electronic technology innovation is one way for patients and nurses to interact more quickly. A component that can calculate, remember, and make choices using a microcontroller is needed to make this device. The microcontroller is a chip or an integrated circuit (IC) that contains a processor and a flash memory capable of reading and writing up to 1000 times. The cost of development is inexpensive because it can be minimized and refilled with other programs as needed. Wireless network technology has been widely researched by previous researchers. The WI-FI Microcontroller based Capacitive Water Wireless Sensor System Model was studied by Suryono [6]. [7] Thakare Obtaining Information About Neighboring Street Lights Using the WIFI Mesh Network was researched by S.N [8]. Riviezzo [9] investigated Wi-Fi Activated Speech Recognition Control Nodes. Xiao [10] investigated the design of Wi-Fi internet of things (IoT)-based household appliance control systems for Smart Homes. An efficient remote control system using SMS and Wi-Fi technology for outdoor security lighting applications was studied by Akorede [11]. Dai [12] investigated a smart car design based on Wi-Fi video capture and OpenCV motion control. A 220-volt power switch controlled through Wi-Fi was studied by Gao [13]. Jarande's research [14]. The Internet based monitoring and protection of the smart PV grid system was studied by Pramono [15]. Aalsalem's Campus Sense Smart Vehicle Parking Monitoring and Management System using ANPR Cameras and Android Phones was researched by Aalsalem [16]. In this study, Sriyanka [17] investigated Intelligent Environmental Monitoring through the Internet of Things (IoT) using Raspberry Pi 3. Occupancy estimation based on environmental sensors in buildings through

IHMMMLR was studied by Chen [18]. Yu [19] has researched a real-time carbon dioxide emission system based on participatory sensing technology based on Yu [19]. Tian-He [20] investigated Comprehensive Monitoring and Analysis Instrument Design for Mine Environment. Smart Community Monitoring Platform for Smart Homes that Actually Work was researched by Nettikadan [21]. The equipment control and environmental environment of the smart home were investigated by Shiqi [22]. An Automated Service Request System for Security in Smart Homes Using the Internet of Things (IoT) was investigated by Madupu [23]. Edge-Based Smart Bura's Parking Solution Using Camera Networks and Deep Learning were researched by Bura [24]. The Monitoring and Ventilation Control System for Multi-story Historic Buildings was studied by Singh [25]. Nurse calling devices can facilitate communication between patients and nurses in the hospital area when patients need aid in an emergency. This nurse calling device is different from the nurse caller that has been researched by previous researchers. This nurse caller uses wireless technology. so that the installation does not require a lot of cables.

II. MATERIALS AND METHODS

2.1. Methods

First, when you press the call button, the Arduino microcontroller receives input from the button call sensor information, which is then processed by the Arduino microcontroller to be transmitted using the (HM-10) Bluetooth module to send a signal, and the signal is received by the receiver. The Bluetooth module (HM-10) is then processed by the Arduino microcontroller client to be displayed on the Liquid Crystal Display (LCD) in the form of a patient call, Light Emitting Diode (LED) and buzzer light up. Secondly, when the stop button is pressed, the Arduino microcontroller receives input from the button stop sensor information, then it is processed by the Arduino microcontroller to be emitted using the Bluetooth module (HM-10) to send a signal, and the signal is received by the (HM-10) Bluetooth module receiver, then processed by the Arduino microcontroller client to be displayed on the LCD in the form of stop call, LED off, and buzzer off [26, 27]. The ATMEGA8 microcontroller is used to process input information from the sensor button and then transmit it to the Bluetooth transmitter. Additionally, the LCD displayed the characters and numbers. The LED is used as an indicator and the buzzer as an alarming unit. When the stop button was pressed, the HM-10 Bluetooth transmitter [28] sent a signal, and it was received by the HM-10 Bluetooth receiver [29]. The characters or numbers on the LCD were deleted, and the LED indicator and buzzer alarm were turned off. From the implemented system, it can be seen that the program starts with register initialization, displays the display and waits for the call and stop button presses. When the call button is pressed, information is sent and received for an active call. When the stop button is pressed, information is sent and received for the deleted call. When the call is active, the LED lights up, and when the call is not active, the LED turns off. There are four buttons to call the nurse and four buttons to stop calling the nurse on the system. The nurse calling device operates on a 220-volt AC power supply and has a current safety. There are four indicator LEDs for each patient bed. When the patient in the bed first presses the nurse call button, the LED indicator for the first patient bed lights up. In addition to the LED indicator, there is a display in the form of an LCD to display patient status. The nurse calling device operates on a 220-volt AC power supply and has a current safety. The design of the power supply circuit used the Proteus application on the computer to create the module. This power supply circuit in this module served as a voltage supply to all circuits that used direct current (DC) voltage. The working principle of the power supply was to change the alternating current (AC) voltage to a DC voltage by using a transformer as a voltage reducer and a diode as a voltage rectifier. In this module, the power supply changed the alternating current (AC) to direct current (DC) by using a regulator integrated circuit (IC) 7805. The 5-volt voltage produced was used to supply the minimum system. The minimum system circuit design uses applications on laptops. The application used in making this module is proteus software. The minimum system circuit in this module functions as the overall module work controller. The workings of the minimum circuit system utilize the storage capacity of an ATMEGA8 integrated circuit (IC). In this example, the ATMEGA8 IC is given a program that will control the module work system as a whole. The application program used in this module is CVAvr. From the design, it can be seen that the LCD display is connected to Port C on the microcontroller. For the receiving module connected with pins TX and RX the led indicator is connected to Port D. The nurse call button and the call stop button are connected to pin C on the microcontroller. Meanwhile, the transmitter module is connected to pin D of the microcontroller. System testing in the analysis of the test design, there are 2 parameters to be tested, namely: (1) Indoor testing: Place the testing device on the horizontal and vertical distances in the room to determine the distance that the wireless module in the room can travel. (2) Outdoors testing device on the horizontal and vertical distance outside the room to know the distance that can be traveled by the wireless module outside the room. Engineering wireless devices to call nurses using one receiver and one transmitter with an HM-10 Bluetooth module. On the transmitter there are four remotes that are connected with a call button and a stop button, so that one remote can be placed in each room/bed.

2.2. Materials

The Transmitter: When the switch is pressed, the voltage from the grid of the state electricity company (PLN) will enter the power supply to change the voltage to direct current. The minimum system circuit gets a 5-volt direct current voltage supply that enters the Bluetooth (HM-10) module, which will be processed for removal at the specified port. In this device, ports D0 and D1 are set as the outputs of the maximum system circuit that will be connected to the Bluetooth

(HM-10) module. The outputs from ports D0 and D1 are used to activate the Bluetooth (HM-10) module, which functions as a wireless information sender. While on port B4, B5, C0, C1, C2, C3, C4, and C5 are set to activate buttons that function as call and stop buttons. The Receiver: When the switch is pressed, the voltage from the grid will enter the power supply to change the voltage to direct current. The minimum system circuit gets a 5-volt direct current voltage supply that enters the Bluetooth (HM-10) module, which will be processed for removal at the specified port. In this device, ports D0 and D1 are set as the outputs of the maximum system circuit that will be connected to the Bluetooth (HM-10) module. The outputs from ports D0 and D1 are used to activate the Bluetooth (HM-10) module, which functions as a receiver of information sent by the sending module. On ports C0, C1, C2, C3, C4, and C5 are set to activate the liquid crystal display (LCD), which functions as a room/bed number viewer that makes calls and stops calls. On port C6, is set as the output of the reset button, which functions to repeat or restart the program from scratch. On port D3, the output of the buzzer is set as an alarm when the call button is pressed. While on ports D4, D5, D6, and D7 are set to activate light emitting diodes (LEDs) that function as room/bed number viewers that make calls as well as LCDs, the difference is the appearance, because the LEDs only expose light, while the LCD displays character letters and numbers.

III. THE CONS OF A WIRELESS NURSE CALL SYSTEM

1. The hardware costs can be higher.
2. Device batteries need to be changed on a regular basis, often every one to two years [33].

IV. The benefits of adopting a wireless nurse call system

1. Instant communication and quicker response time The wireless nurse call systems give patients access to timely care, especially in terrible situations where response time decides the odds in the battle of life and death. The wireless nurse call system alerts the medical staff and sends them the exact location of the patient that requires or needs medical attention. If the patient is wearing a wireless hand transmitter, then she or he can be found anywhere inside the care facility or hospital. an advanced mobile application for smartphones, which enables them to be informed and react even when they are not in the vicinity of the nurse's station. The iNurse mobile application is an advanced replacement for pagers or DECT phones. How so? Because it also enables nurses to see which personnel are currently working and allows them to make direct assistance calls without using a SIM card. 2. Relieving the personnel Many hospitals still rely on paper documentation, which delays the process of patient admission, medical research, and diagnosis. Nevertheless, healthcare management can be optimized. With an intelligent wireless nurse call system, you can digitalize your healthcare administration and save all information inside a cloud database connected to the hospital network. That is not all! An IP nurse call system like Nurse Care allows medical personnel to store and access patient information instantly. The Nurse Tab, an interactive room touchscreen display, and the interactive corridor display Info Tab connect the nurse call system, the entire healthcare management system, and nursing documentation into one device. Nurses can store patient information and updates from inside a patient's room or the hallway. The latter eradicates the need to fill out documents, and it likewise prevents the loss of data or repetition of information. In most cases, this happens when the nurses transcribe handwritten data into electronic form. 3. Better workflow and fewer disturbances Enhanced workflow cuts down on any potential losses, both in time, money, and knowledge. With an intelligent nurse call system, every piece of data is stored instantly, accessible with a simple touch, and transmitted among medical staff. Better communication leads to better decisions and better organization. It likewise helps lower the number of medical errors, which affects not only the hospital's budget but also its reputation. With nurse care, there are fewer disturbances in the healthcare process since it provides prompt and timely patient care. It helps nurses stay organized and keeps them focused on dealing with patient care, not distractions. 4. Fewer mistakes and cost-effectiveness Monitoring technology benefits both the medical staff and the hospital management. Healthcare workers treat several patients throughout the day. In this process, data gets handed down many times, resulting in loss of data or even a medical error. With monitoring technology, such as Nurse Tab, healthcare processes are under constant supervision, guaranteeing patient safety. The Nurse Tab enables medical personnel to log (a) The personnel assigned to work in a specific patient room, (b) The time they will arrive and the time allotted (c). Nursing notes, (d) Services provided, (e) The patient's name in care. When you consider the long-term benefits and the initial cost, you realize that employing a wireless nurse call system pays off. Especially when you can connect it to your existing nurse call system and use the installation that is already in the building. You can personalize and select only those features that will benefit you the most. 5. Satisfied staff, satisfied patients Satisfactory, attentive, and efficient patient care results in pleased and comfortable patients, who are likely to recover quicker. Wireless nurse call systems quicken the treatment processes, optimize workflow, and decrease the level of stress in hospitals and other caring facilities. These intelligent solutions are easy to use, which helps keep hospital tasks moving smoothly and efficiently [31].

V. CHALLENGES OF ADOPTING THE NURSE CALLING SYSTEMS

Adopting a wireless nurse call system presents several challenges, including potential signal interference, battery maintenance, and the possibility of single points of failure. Additionally, staff resistance to change, budget constraints, and compatibility issues with existing systems can hinder implementation.

Here's a more detailed look at the challenges:

1. **Technical Challenges:**

- **Signal Interference:**

Wireless systems can be susceptible to interference from other devices and systems within the facility, potentially causing unreliable calls.

- **Battery Maintenance:**

Wireless devices require regular battery changes, which can add to the maintenance workload and potentially lead to system downtime.

- **Single Point of Failure:**

If a critical component fails in a wireless system, the entire system may be affected, requiring a complete shutdown and potential disruption to patient care.

- **Hardware Costs:**

While wireless systems offer flexibility, the hardware costs can be higher than wired systems.

- **Dome Lights:**

Even in a wireless system, dome lights in hallways may still require wiring for power, adding to the installation complexity.

2. **Organizational and Implementation Challenges:**

- **Staff Resistance:**

Nurses and other healthcare professionals may resist adopting new technology, fearing that it will disrupt their workflows or increase their workload.

- **Budget Constraints:**

Implementing a new nurse call system can be expensive, and hospital administrators may not have the budget for installation, equipment, and ongoing maintenance.

- **Compatibility Issues:**

Many hospitals have existing communication systems, and new wireless systems may not be compatible, requiring additional costs and delays for integration.

- **Patient Needs:**

Not all patients have the same needs, and some may have difficulty using the system due to hearing or speech impairments, requiring alternative solutions.

- **Implementation and Training:**

A successful implementation requires adequate training for staff on how to use the new system, which can be a time-consuming and challenging process.

Other Considerations:

- **Logistics and Disruption:**

Implementing a new system can disrupt normal operations, especially in existing buildings where changes may require extensive planning and coordination, as seen in a case study of a rural Scottish community.

- **Logistical Challenges:**

Issues like transportation disruption and hotel accommodation unavailability can further complicate the implementation process, as highlighted in a case study.

- **Need for Pre-Programming:**

In some cases, pre-programming the system before installation can help streamline the process and ensure it meets specific requirements, as demonstrated in another case study.

- **Need for User Input:**

Gathering feedback from users, including patients and healthcare providers, throughout the implementation process is crucial for ensuring that the system meets their needs and is effectively used.

CONCLUSION

After designing, experimenting, testing, and collecting data, the author comes to the conclusion that using a Bluetooth module to create a nurse call device will make installation easier and produce a neater final product. To build these devices, we needed a component that could compute, remember, and make decisions using a microcontroller. If the orders detected on the LCD and LED are displayed, it will be easier for nurses to learn about the patient's room that is making the call. Using both horizontal and vertical measurements outside the room, we show that wind can change the frequency of the Bluetooth module, leading to different distance estimations. The consequences of setting up the wireless nurse calling system are discussed [32]. The disadvantages of the nurse call system are mentioned [34].

REFERENCES

1. F. Vannieuwenborg et al., "Techno-economic evaluation of an ontology-based nurse call system via discrete event simulations," in 2014 IEEE 16th International Conference on e Health Networking, Applications and Services (Healthcom), 2014, no. Ssh, pp. 82–87.
2. C. Sharma and D. K. Gautam, "Design development and implementation of wired Nurse calling system," in 2015 International Conference on Green Computing and Internet of Things (ICGCIoT), 2015, vol. 9, no. 5, pp. 1258–1262.
3. T. Maekawa et al., "A study on automatic evaluation method of pointing and calling for nurse education," in 2015 International Conference on Intelligent Informatics and Biomedical Sciences (ICIIBMS), 2015, pp. 383–384.
4. N. Khera, S. Tiwari, R. P. Singh, T. Ghosh, and P. Kumar, "Development of android based smart home and nurse calling system for differently abled," in 2016 5th International Conference on Wireless Networks and Embedded Systems (WECON), 2016, pp. 1–4.
5. M. S. Mahmud, M. A. Majumder, A. K. Tushar, M. M. Kamal, A. Ashiquzzaman, and M. R. Islam, "Real-time feedback-centric nurse calling system with archive monitoring using Raspberry Pi," in 2017 4th International Conference on Networking, Systems and Security (NSysS), 2017, vol. 2018-Janua, pp. 1–5.
6. S. Suryono, W. Widowati, S. P. Putro, and S. Sunarno, "A Capacitive Model of Water Salinity Wireless Sensor System Based on WIFIMicrocontroller," in 2018 6th International Conference on Information and Communication Technology (ICoICT), 2018, vol. 0, no. c, pp. 211–215.
7. S. Thakare and P. H. Bhagat, "Arduino-Based Smart Irrigation Using Sensors and ESP8266 WiFi Module," in 2018 Second International Conference on Intelligent Computing and Control Systems (ICICCS), 2018, no. Iccics, pp. 1–5.
8. C. S.N., S. Singha, S. Ghorai, N. V., and B. Samuel, "Getting Information about the Neighbour Street Light Using WIFI Mesh Network," in 2018 International Conference on Design Innovations for 3Cs Compute Communicate Control (ICDI3C), 2018, pp. 130–132.
9. N. Riviezzo and B. Martinez, "WiFi Enabled Speech Recognition Controller Node," in 2017 IEEE Long Island Systems, Applications and Technology Conference (LISAT), 2017, pp. 1–6.
10. Z. Xiao, D. Liu, D. Cao, and X. Wang, "Design of Home Appliance Control System in Smart Home based on WiFi IoT," in 2018 IEEE 3rd Advanced Information Technology, Electronic and Automation Control Conference (IAEAC), 2018, no. Iaeac, pp. 765–770.
11. M. F. Akorede, J. J. Fatigun, J. A. Opaluwa, and E. Pouresmaeil, "Efficient remote control system using SMS and WiFi technology for outdoor security lighting applications," in 2017 IEEE PES PowerAfrica, 2017, pp. 512–517.
12. G. Dai and P. Wang, "Design of intelligent car based on WiFi video capture and OpenCV gesture control," in 2017 Chinese Automation Congress (CAC), 2017, vol. 112, no. 483, pp. 4103–4107.
13. X. Gao, B. Zhang, and S. Li, "A 220-volts power switch controlled through WiFi," in 2016 First IEEE International Conference on Computer Communication and the Internet (ICCCI), 2016, pp. 526–529.
14. P. B. Jarande, S. P. Murakar, N. S. Vast, N. P. Ubale, and S. S. Saraf, "Robotic Vacuum Cleaner Using Arduino with Wifi," in 2018 Second International Conference on Inventive Communication and Computational Technologies (ICICCT), 2018, no. 9, pp. 1513–1517.
15. S. H. Pramono, S. N. Sari, and E. Maulana, "Internet-based monitoring and protection on PV smart grid system," in 2017 International Conference on Sustainable Information Engineering and Technology (SIET), 2017, vol. 2018-Janua, pp. 448–453.
16. M. Y. Aalsalem and W. Z. Khan, "CampusSense - A Smart Vehicle Parking Monitoring and Management System using ANPR Cameras and Android Phones," vol. 5, no. 2, 20170.
17. Sriyanka and S. R. Patil, "Smart Environmental Monitoring through Internet of Things (IoT) using RaspberryPi 3," in 2017 International Conference on Current Trends in Computer, Electrical, Electronics and Communication (CTCEEC), 2017, pp. 595–600.
18. Z. Chen, Q. Zhu, M. K. Masood, and Y. C. Soh, "Environmental Sensors-Based Occupancy Estimation in Buildings via IHMM-MLR," IEEE Trans. Ind. Informatics, vol. 13, no. 5, pp. 2184–2193, Oct. 2017.
19. R. Yu, W. Wu, N. Xia, H. Geng, and M. Liu, "Real-time carbon dioxide emission monitoring system based on participatory sensing technology," in The Fourth International Workshop on Advanced Computational Intelligence, 2011, pp. 230–235.
20. W. Tian-He, M. Li, W. Zhong-Hua, Y. Lin, W. Can, and S. Qin-Peng, "Design of Comprehensive Monitoring and Analysis Instrument for Mine Environment," in 2018 Chinese Automation Congress (CAC), 2018, pp. 3885–3889.
21. D. Nettikadan and R. M. S. Subodh, "IOT Based Smart Community Monitoring Platform for Custom Designed Smart Homes," in 2018 International Conference on Current Trends towards Converging Technologies (ICCTCT), 2018, pp. 1–6.
22. Z. Shiqi, W. Xiaohui, and C. Hongbing, "Equipment control and environmental monitoring design of smart home," in 2018 Chinese Control and Decision Conference (CCDC), 2018, pp. 513–517.
23. P. K. Madupu and B. Karthikeyan, "Automatic Service Request System for Security in Smart Home Using IoT," in 2018 Second International Conference on Electronics, Communication and Aerospace Technology (ICECA), 2018, no. Iceca, pp. 1413–1418.

24. H. Bura, N. Lin, N. Kumar, S. Malekar, S. Nagaraj, and K. Liu, "An Edge Based Smart Parking Solution Using Camera Networks and Deep Learning," in 2018 IEEE International Conference on Cognitive Computing (ICCC), 2018, pp. 17–24.
25. A. Singh, Y. Pandey, A. Kumar, M. K. Singh, A. Kumar, and S. C. Mukhopadhyay, "Ventilation Monitoring and Control System for High Rise Historical Buildings," IEEE Sens. J., vol. 17, no. 22, pp. 7533–7541, Nov. 2017.
26. S. Widadi, M. K. Huda, I. Ahmad, and O. Tanane, "Atmega328P-based X-ray Machine Exposure Time Measurement Device with an Android Interface," J. Robot. Control, vol. 1, no. 3, pp. 81–85, 2020.
27. A. Hassan et al., "A Wirelessly Controlled Robot-based Smart Irrigation System by Exploiting Arduino," J. Robot. Control, vol. 2, no. 1, pp. 29–34, 2020.
28. K. Kunal, A. Z. Arfianto, J. E. Poetro, F. Waseel, and R. A. Atmoko, "Accelerometer Implementation as Feedback on 5 Degree of Freedom Arm Robot," J. Robot. Control, vol. 1, no. 1, pp. 31–34, 2020.
29. J. Crha, O. Tupa, J. Mares, and A. Prochazka, "Navigation of robotic platform for gait disorders monitoring," in 2016 International Conference on Applied Electronics (AE), 2016, vol. 2016-Sept, pp. 57–60.
30. S. Widadi, S. Al B. Munir, N. Shahu, I. Ahmad, I. Al Barazanchi, "Automatic Wireless Nurse Caller", Journal of Robotics and Control (JRC) Vol. 2, No. 5, pp. 380-384, DOI: 10.18196/jrc.25111, September 2021.
31. <https://caretronic.com/benefits-wireless-nurse-call-system>.
32. A. H. Muhammad, A. Y. Abdullahi, A. Abba, A. Isah, A. A. Yako, M. A. Baballe, "The Benefits of Adopting a Wireless Nurse Call System", Global Journal of Research in Medical Sciences, Volume 02, Issue <https://gjrpublication.com/gjrms/>. 03, May-June 2022, Journal homepage:
33. https://www.google.com/search?q=disadvantages+of+a+Wireless+Nurse+Call+System&sca_esv=179241790b8982d2&ei=uYvkZb3GDsGzhbIP5yd6A0&ved=0ahUKEwj9ho2oqdiEAXBWUEAHWd2B90Q4dUDCBA&uact=5&oq=disadvantages+of+a+Wireless+Nurse+Call+System&gs_lp=Egxnd3Mtd2l6LXNlcnAiLWRpc2FkdmdFudGFnZXMgb2YgYSBXaXJlbGVzcyB0dXJzZSBdYWxsIFN5c3RlbTIIeAAAYiQUYogQyCBAAGIAEGKIEMggQABiABBiiBEiu3wFQ3Z4BWLizAXACeACQAQCYAYkCoAGgEaoBAzItObgBA8gBAPgBAZgCBaAC0gXCAG4QABiABBiKBRiGAXiwA8ICCBahGKABGMMEmAMaAiAYBkAYBkgcFMi4wLjM&scient=gws-wiz-serp#ip=1.
34. M. D. Mohammed, M. A. Said, M. A. Baballe, "THE DRAWBACKS OF A WIRELESS MEDICAL ALERT SYSTEM", TMP UNIVERSAL JOURNAL OF RESEARCH AND REVIEW ARCHIVES VOLUME 3 | ISSUE 1 | YEAR 2024 | JAN – MAR 2024.

CITATION

Sarafadeen L. L., Aliyu S., Livingston S. A., & Muhammad A. B. (2025). The Challenges of Adopting a Wireless Nurse call system. In Global Journal of Research in Engineering & Computer Sciences (Vol. 5, Number 3, pp. 36–41).

<https://doi.org/10.5281/zenodo.15388093>